

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106



Date: 11/02/2020

OPERATOR NAME:
DESMOND L LYONS
Permit #: 35876
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:28
INTERNAL STANDARD	VERIFIED	02:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:29
BLANK TEST	0.00	02:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:30
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	INTERFERENCE	02:32

Interference Detected



SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/02/2020

OPERATOR NAME:
DESMOND L LYONS
Permit #: 35876
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:39
INTERNAL STANDARD	VERIFIED	02:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:42
BLANK TEST	0.00	02:43

