

SUPERVISOR TEST

GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 11/23/2020

OPERATOR NAME:
 ROBERT M FOO
 Permit #: 46059
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:47
INTERNAL STANDARD	VERIFIED	01:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:50
BLANK TEST	0.00	01:51

