

SUPERVISOR TEST



GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/30/2020

OPERATOR NAME:
NICHOLAS K RYSEDORPH
Permit #: 48769
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:29
INTERNAL STANDARD	VERIFIED	02:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:30
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:31
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:33
BLANK TEST	0.00	02:33



SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106



Date: 11/30/2020

OPERATOR NAME:
NICHOLAS K RYSEDORPH
Permit #: 48769
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:25
INTERNAL STANDARD	VERIFIED	02:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:27

Blank Error



SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/30/2020

OPERATOR NAME:
NICHOLAS K RYSEDORPH
Permit #: 48769
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:16
INTERNAL STANDARD	VERIFIED	02:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:17

Blank Error

