

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/07/2020

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:04
INTERNAL STANDARD	VERIFIED	00:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:05
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:06
BLANK TEST	0.00	00:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:07
BLANK TEST	0.00	00:08

