

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/21/2020

OPERATOR NAME:
DESMOND L LYONS
Permit #: 35876
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:28
INTERNAL STANDARD	VERIFIED	01:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:28
BLANK TEST	0.00	01:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:29
BLANK TEST	0.00	01:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:31
BLANK TEST	0.00	01:31

