

SUPERVISOR TEST

-----  
 GLENS FALLS CITY  
 POLICE DEPT  
 DATAMASTER dmt: 117106  
 -----

Date: 09/28/2020

OPERATOR NAME:  
 NICHOLAS K RYSEDORPH  
 Permit #: 40769  
 Operator Agency/Dept: GLENS FALLS CITY PD  
 Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:20
INTERNAL STANDARD	VERIFIED	03:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:20
BLANK TEST	0.00	03:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:21
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:23
BLANK TEST	0.00	03:23

