

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/07/2020

OPERATOR NAME:
DESMOND L LYONS
Permit #: 35786
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:56
INTERNAL STANDARD	VERIFIED	01:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:56
BLANK TEST	0.00	01:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:58
BLANK TEST	0.00	01:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:59
BLANK TEST	0.00	02:00

