

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/11/2021

OPERATOR NAME:
JOHN F HOGAN
Permit #: 49269
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:10
INTERNAL STANDARD	VERIFIED	04:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:11
BLANK TEST	0.00	04:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	04:12
BLANK TEST	0.00	04:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:14
BLANK TEST	0.00	04:14

