

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 01/18/2021

OPERATOR NAME:  
GERALD C WILLETTE  
Permit #: 32550  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:28
INTERNAL STANDARD	VERIFIED	02:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:29
BLANK TEST	0.00	02:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:30
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:31
BLANK TEST	0.00	02:32

