

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 01/25/2021

OPERATOR NAME:  
JOHN F HOGAN  
Permit #: 49269  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:05
INTERNAL STANDARD	VERIFIED	02:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:08
BLANK TEST	0.00	02:09

