

SOLUTION CHANGE



GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/01/2021

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:47
INTERNAL STANDARD	VERIFIED	03:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:48
BLANK TEST	0.00	03:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:49
BLANK TEST	0.00	03:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:50
BLANK TEST	0.00	03:51



SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/01/2021

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:03
INTERNAL STANDARD	VERIFIED	03:03
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	03:04
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:05
BLANK TEST	0.00	03:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:06
BLANK TEST	0.00	03:07

