

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/08/2021

OPERATOR NAME:
TYLER R MELLO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:34
INTERNAL STANDARD	VERIFIED	06:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:35
BLANK TEST	0.00	06:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:36
BLANK TEST	0.00	06:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	06:37
BLANK TEST	0.00	06:38

