

SUPERVISOR TEST

-----  
GLEN FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 02/15/2021

OPERATOR NAME:  
JOHN F HOGAN  
Permit #: 49269  
Operator Agency/Dept: GLEN FALLS CITY PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:13
INTERNAL STANDARD	VERIFIED	00:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:13
BLANK TEST	0.00	00:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:15
BLANK TEST	0.00	00:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:16
BLANK TEST	0.00	00:17

