

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/01/2021

OPERATOR NAME:
SHAWN A LOVELACE
Permit #: 30657
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:13
INTERNAL STANDARD	VERIFIED	01:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:15
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:17

