

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 03/15/2021

OPERATOR NAME:  
TYLER R MELLO  
Permit #: 46781  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:59
INTERNAL STANDARD	VERIFIED	03:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:59
BLANK TEST	0.00	04:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:00
BLANK TEST	0.00	04:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:02
BLANK TEST	0.00	04:03

