

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 05/03/2021

OPERATOR NAME:  
GERALD C WILLETTE  
Permit #: 32550  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:23
INTERNAL STANDARD	VERIFIED	06:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:24
BLANK TEST	0.00	06:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:25
BLANK TEST	0.00	06:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	06:26
BLANK TEST	0.00	06:27

