

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/05/2021

OPERATOR NAME:
SHAWN A LOVELACE
Permit #: 30657
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:31
INTERNAL STANDARD	VERIFIED	14:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:32
BLANK TEST	0.00	14:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:33
BLANK TEST	0.00	14:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:34
BLANK TEST	0.00	14:35

