

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/10/2021

OPERATOR NAME:
CRAIG M RAGGI
Permit #: 49219
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:41
INTERNAL STANDARD	VERIFIED	01:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:41
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:43
BLANK TEST	0.00	01:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:44
BLANK TEST	0.00	01:45

