

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/17/2021

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:16
INTERNAL STANDARD	VERIFIED	02:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:17
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:19
BLANK TEST	0.00	02:20

