

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/21/2021

OPERATOR NAME:
TYLER R MELLO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:24
INTERNAL STANDARD	VERIFIED	02:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:24
BLANK TEST	0.00	02:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:25
BLANK TEST	0.00	02:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:27
BLANK TEST	0.00	02:28

