

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 07/01/2021

OPERATOR NAME:
 DESMOND L LYONS
 Permit #: 35876
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:01
INTERNAL STANDARD	VERIFIED	08:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:02
BLANK TEST	0.00	08:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:03
BLANK TEST	0.00	08:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:04
BLANK TEST	0.00	08:05

