

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/05/2021

OPERATOR NAME:
TYLER R MELLO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:06
INTERNAL STANDARD	VERIFIED	00:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:07
BLANK TEST	0.00	00:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:08
BLANK TEST	0.00	00:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:09
BLANK TEST	0.00	00:10

