

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/12/2021

OPERATOR NAME:
JOHN F HOGAN
Permit #: 49269
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:26
INTERNAL STANDARD	VERIFIED	00:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:27
BLANK TEST	0.00	00:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:28
BLANK TEST	0.00	00:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:29
BLANK TEST	0.00	00:30

