

SOLUTION CHANGE

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 08/02/2021

OPERATOR NAME:  
TYLER R MELLO  
Permit #: 21170  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 46781

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:41
INTERNAL STANDARD	VERIFIED	01:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:41
BLANK TEST	0.00	01:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:43
BLANK TEST	0.00	01:44

