

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/16/2021

OPERATOR NAME:
TYLER R MELLO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:28
INTERNAL STANDARD	VERIFIED	04:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:28
BLANK TEST	0.00	04:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:29
BLANK TEST	0.00	04:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	04:31
BLANK TEST	0.00	04:32

