

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 09/06/2021

OPERATOR NAME:  
JOHN F HOGAN  
Permit #: 49269  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:37
INTERNAL STANDARD	VERIFIED	00:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:38
BLANK TEST	0.00	00:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:39
BLANK TEST	0.00	00:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:40
BLANK TEST	0.00	00:41

