

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 09/13/2021

OPERATOR NAME:  
TYLER R MELLO  
Permit #: 46781  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:08
INTERNAL STANDARD	VERIFIED	02:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:09
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:10
BLANK TEST	0.00	02:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:11
BLANK TEST	0.00	02:12

