

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/20/2021

OPERATOR NAME:
CRAIG M RAGGI
Permit #: 49219
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:13
INTERNAL STANDARD	VERIFIED	01:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:13
BLANK TEST	0.00	01:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:14
BLANK TEST	0.00	01:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:16

