

SOLUTION CHANGE



GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/01/2021

OPERATOR NAME:
MARC J BARRETT
Permit #: 46725
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:26
INTERNAL STANDARD	VERIFIED	01:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:27
BLANK TEST	0.00	01:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:28
BLANK TEST	0.00	01:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:29
BLANK TEST	0.00	01:30



SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/29/2021

OPERATOR NAME:
JOHN F HOGAN
Permit #: 49269
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:42
INTERNAL STANDARD	VERIFIED	01:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:42
BLANK TEST	0.00	01:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:44
BLANK TEST	0.00	01:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:45
BLANK TEST	0.00	01:46

