

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/13/2021

OPERATOR NAME:
JOHN F HOGAN
Permit #: 49269
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:54
INTERNAL STANDARD	VERIFIED	04:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:55
BLANK TEST	0.00	04:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:56
BLANK TEST	0.00	04:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:57
BLANK TEST	0.00	04:58

