

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/20/2021

OPERATOR NAME:
JOHN F HOGAN
Permit #: 49269
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:57
INTERNAL STANDARD	VERIFIED	00:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:57
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:59
BLANK TEST	0.00	01:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:00
BLANK TEST	0.00	01:01

