

SOLUTION CHANGE

-----  
 GLENS FALLS CITY  
 POLICE DEPT  
 DATAMASTER dmt: 117106  
 -----

Date: 01/01/2022

OPERATOR NAME:  
 RYAN S PEDONE  
 Permit #: 35878  
 Operator Agency/Dept: GLENS FALLS CITY PD  
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:10
INTERNAL STANDARD	VERIFIED	06:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:10
BLANK TEST	0.00	06:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:11
BLANK TEST	0.00	06:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:13
BLANK TEST	0.00	06:14

