

SUPERVISOR TEST

GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 01/17/2022

OPERATOR NAME:
 CRAIG M RAGGI
 Permit #: 49219
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 21390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:50
INTERNAL STANDARD	VERIFIED	00:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:51
BLANK TEST	0.00	00:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:52
BLANK TEST	0.00	00:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:53
BLANK TEST	0.00	00:54

