

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/07/2022

OPERATOR NAME:
CRAIG M RAGGI
Permit #: 49219
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:13
INTERNAL STANDARD	VERIFIED	02:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:13
BLANK TEST	0.00	02:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:14
BLANK TEST	0.00	02:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:16
BLANK TEST	0.00	02:16

