

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/28/2022

OPERATOR NAME:
TYLER R MELLO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:46
INTERNAL STANDARD	VERIFIED	05:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:46
BLANK TEST	0.00	05:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:47
BLANK TEST	0.00	05:48
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:49
BLANK TEST	0.00	05:50

