

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/07/2022

OPERATOR NAME:
CRAIG M RAGGI
Permit #: 49219
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:25
INTERNAL STANDARD	VERIFIED	03:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:26
BLANK TEST	0.00	03:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:27
BLANK TEST	0.00	03:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:28
BLANK TEST	0.00	03:29

