

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 03/14/2022

OPERATOR NAME:
 RYAN S PEDONE
 Permit #: 35878
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:45
INTERNAL STANDARD	VERIFIED	03:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:45
BLANK TEST	0.00	03:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:46
BLANK TEST	0.00	03:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:47
BLANK TEST	0.00	03:48

