

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/21/2022

OPERATOR NAME:
TYLER R MELLIO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:03
INTERNAL STANDARD	VERIFIED	02:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:04
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07

