

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 04/04/2022

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:46
INTERNAL STANDARD	VERIFIED	03:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:46
BLANK TEST	0.00	03:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:47
BLANK TEST	0.00	03:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:48
BLANK TEST	0.00	03:49

