

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/11/2022

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:00
INTERNAL STANDARD	VERIFIED	08:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:00
BLANK TEST	0.00	08:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:01
BLANK TEST	0.00	08:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	08:03
BLANK TEST	0.00	08:04

