

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/18/2022

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:34
INTERNAL STANDARD	VERIFIED	05:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:35
BLANK TEST	0.00	05:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:36
BLANK TEST	0.00	05:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:37
BLANK TEST	0.00	05:38

