

SUPERVISOR TEST

-----  
 GLENS FALLS CITY  
 POLICE DEPT  
 DATAMASTER dmt: 117106  
 -----

Date: 04/25/2022

OPERATOR NAME:  
 KIRSTEN M LUNDER  
 Permit #: 41111  
 Operator Agency/Dept: GLENS FALLS CITY PD  
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:11
INTERNAL STANDARD	VERIFIED	05:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:11
BLANK TEST	0.00	05:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:13
BLANK TEST	0.00	05:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	05:14
BLANK TEST	0.00	05:15

