

SOLUTION CHANGE




GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/01/2022

OPERATOR NAME:
TYLER R MELLO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:20
INTERNAL STANDARD	VERIFIED	06:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:20
BLANK TEST	0.00	06:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:21
BLANK TEST	0.00	06:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	06:23
BLANK TEST	0.00	06:24



SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/02/2022

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:22
INTERNAL STANDARD	VERIFIED	00:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:23
BLANK TEST	0.00	00:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:24
BLANK TEST	0.00	00:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:25
BLANK TEST	0.00	00:26

