

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/27/2022

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:31 |
| INTERNAL STANDARD | VERIFIED | 00:31 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 00:32 |
| BLANK TEST | 0.00 | 00:32 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.09 | 00:33 |
| BLANK TEST | 0.00 | 00:34 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.09 | 00:34 |
| BLANK TEST | 0.00 | 00:35 |

