

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/04/2022

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:33
INTERNAL STANDARD	VERIFIED	03:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:33
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:34
BLANK TEST	0.00	03:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:36
BLANK TEST	0.00	03:37

