

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 07/18/2022

OPERATOR NAME:
 ROBERT M FOO
 Permit #: 46059
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 01:30 |
| INTERNAL STANDARD | VERIFIED | 01:30 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 01:31 |
| BLANK TEST | 0.00 | 01:31 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 01:32 |
| BLANK TEST | 0.00 | 01:33 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 01:33 |
| BLANK TEST | 0.00 | 01:34 |

