

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 07/25/2022

OPERATOR NAME:  
KIRSTEN M LUNDER  
Permit #: 44441  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 22100

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 04:37 |
| INTERNAL STANDARD       | VERIFIED | 04:37 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 04:38 |
| BLANK TEST              | 0.00     | 04:39 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 04:39 |
| BLANK TEST              | 0.00     | 04:40 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.09     | 04:40 |
| BLANK TEST              | 0.00     | 04:41 |

