

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/01/2022

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:46
INTERNAL STANDARD	VERIFIED	02:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:46
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:47
BLANK TEST	0.00	02:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:49
BLANK TEST	0.00	02:49

