

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/29/2022

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:35
INTERNAL STANDARD	VERIFIED	02:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:36
BLANK TEST	0.00	02:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:37
BLANK TEST	0.00	02:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:38
BLANK TEST	0.00	02:39

