

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 09/01/2022

OPERATOR NAME:
 TYLER R MELLO
 Permit #: 46781
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:48
INTERNAL STANDARD	VERIFIED	04:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:49
BLANK TEST	0.00	04:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:50
BLANK TEST	0.00	04:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	04:51
BLANK TEST	0.00	04:52

